

General Dihydrotestosterone (DHT) ELISA Kit | DHT elisa kit

General Dihydrotestosterone (DHT) ELISA Kit

CATALOG #
AAA63820

REACTIVITY
General

PRODUCT SYNONYMS

Dihydrotestosterone (DHT); General Dihydrotestosterone (DHT) ELISA Kit; 5 α -DHT; 5 β -Dihydrotestosterone; Androstanolone; Stanolone; DHT elisa kit

PRODUCT SPECIFICATIONS

Reactivity	General
Specificity	This assay has high sensitivity and excellent specificity for detection of DDAVP. No significant cross-reactivity or interference between DDAVP and analogues was observed.
Samples	Serum, Plasma, Tissue Homogenates, Cell Lysates, Cell Culture Supernates And Other Biological Fluids
Assay Type	Quantitative Competitive
Detection Range	246.9-20,000pg/mL
Sensitivity	99.5pg/mL.
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level DDAVP were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level DDAVP were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kit, all reagents should be kept according to the labels on vials. The TMB Substrate, Wash Buffer, Stop Solution should be stored at 4 degree C. All others should be stored at -20 degree C.

ADDITIONAL INFORMATION

Related Product Information for DHT elisa kit

Intended Uses: The kit is a competitive inhibition enzyme immunoassay technique for the in vitro quantitative measurement of DDAVP in serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids.

Principle of the Assay: This assay employs the competitive inhibition enzyme immunoassay technique. A monoclonal antibody specific to DDAVP has been pre-coated onto a microplate. A competitive inhibition reaction is launched between biotin labeled DDAVP and unlabeled DDAVP (Standards or samples) with the pre-coated antibody specific to DDAVP. After incubation the unbound conjugate is washed off. Next, avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. The amount of bound HRP conjugate is reverse proportional to the concentration of DDAVP in the sample. After addition of the substrate solution, the intensity of color developed is reverse proportional to the concentration of DDAVP in the sample.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.