

Porcine Complement Component 5a (C5a) ELISA Kit | C5a elisa kit

Porcine Complement Component 5a (C5a) ELISA Kit

CATALOG #
AAA65842

REACTIVITY
Porcine


PRODUCT SYNONYMS

Complement Component 5a (C5a); Porcine Complement Component 5a (C5a) ELISA Kit; C5a elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Porcine
Specificity	This assay has high sensitivity and excellent specificity for detection of C5a. No significant cross-reactivity or interference between C5a and analogues was observed.
Samples	Serum, plasma and other biological fluids.
Assay Type	Sandwich
Detection Range	0.156-10ng/mL
Sensitivity	0.072ng/mL
Intra-assay Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level C5a were tested 20 times on one plate, respectively. Intra-Assay: CV<10%
Inter-assay Precision	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level C5a were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100. Inter-Assay: CV<12%
Preparation and Storage	For unopened kit, all reagents should be kept according to the labels on vials. The TMB Substrate, Wash Buffer, Stop Solution should be stored at 4 degree C. All others should be stored at -20 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for C5a elisa kit	Intended Uses: The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of C5a in porcine serum, plasma and other biological fluids. Principle of the Assay: The microtiter plate provided in this kit has been pre-coated with an antibody specific to C5a. Standards or samples are then added to the appropriate microtiter plate wells with a biotin-conjugated antibody preparation specific to C5a. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain C5a, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm +/- 10nm. The concentration of C5a in the samples is then determined by comparing the O.D. of the samples to the standard curve.
---	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/porcine-elisa-kits/complement-component-5a-c5a/65842>