

Monkey thyroid stimulating hormone (TSH) ELISA Kit | TSH elisa kit

Monkey thyroid stimulating hormone, TSH ELISA Kit

CATALOG #
AAA81000

REACTIVITY
Monkey


PRODUCT SYNONYMS

thyroid stimulating hormone (TSH); Monkey thyroid stimulating hormone; TSH ELISA Kit; Monkey thyroid stimulating hormone (TSH) ELISA kit; TSH elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Monkey
Specificity	This assay has high sensitivity and excellent specificity for detection of monkey TSH. No significant cross-reactivity or interference between monkey TSH and analogues was observed.
Samples	Serum, plasma, tissue homogenates
Assay Type	Quantitative Sandwich
Detection Range	0.3 uIU/ml-12 uIU/ml
Sensitivity	<0.15 uIU/ml
Intra-assay Precision	Intra-assay Precision (Precision within an assay): CV%<15%. Three samples of known concentration were tested twenty times on one plate to assess.
Inter-assay Precision	Inter-assay Precision (Precision between assays): CV%<15%. Three samples of known concentration were tested in twenty assays to assess.
Preparation and Storage	Unopened test kits should be stored at 2 to 8 degree C upon receipt. Please refer to pdf manual for further storage instructions.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for TSH elisa kit

Principle of the Assay: This assay employs the quantitative sandwich enzyme immunoassay technique. Antibody specific for TSH has been pre-coated onto a microplate. Standards and samples are pipetted into the wells with a Horseradish Peroxidase (HRP) conjugated antibody specific for TSH. Following a wash to remove any unbound reagent, a substrate solution is added to the wells and color develops in proportion to the amount of TSH bound in the initial step. The color development is stopped and the intensity of the color is measured.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

