

Rabbit Methemoglobin ELISA Kit | MHB elisa kit

Rabbit Methemoglobin ELISA Kit

CATALOG #
AAA82047

REACTIVITY
Rabbit


PRODUCT SYNONYMS

Methemoglobin; Rabbit Methemoglobin ELISA Kit; MHB elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Specificity	This assay has high sensitivity and excellent specificity for detection of MHB. No significant cross-reactivity or interference between MHB and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between MHB and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive
Sensitivity	0.1 umol/L
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for MHB elisa kit	Intended Uses: This MHB ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Rabbit MHB. This ELISA kit for research use only, not for therapeutic or test applications! Principle of the Assay: MHB ELISA kit applies the competitive enzyme immunoassay technique utilizing an anti-MHB antibody and an MHB-HRP conjugate. The assay sample and buffer are incubated together with MHB-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the MHB concentration since MHB from samples and MHB-HRP conjugate compete for the anti-MHB antibody binding site. Since the number of sites is limited, as more sites are occupied by MHB from the sample, fewer sites are left to bind MHB-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The MHB concentration in each sample is interpolated from this standard curve.
---	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/mhb-rabbit-elisa-kits/methemoglobin/82047>