

Rabbit Fibrinogen Degradation Product+B74 ELISA Kit | FDP elisa kit

Rabbit Fibrinogen Degradation Product+B74 ELISA Kit

CATALOG #
AAA83950

REACTIVITY
Rabbit


PRODUCT SYNONYMS

Fibrinogen Degradation Product+B74; Rabbit Fibrinogen Degradation Product+B74 ELISA Kit; FDP elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Rabbit
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Competitive or Sandwich
Detection Range	1.0-25ug/mL
Sensitivity	0.1ug/mL
Intended Uses	This FDP ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Rabbit FDP. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



ADDITIONAL INFORMATION

Related Product Information for FDP elisa kit

Principle of the Assay: FDP ELISA kit applies the quantitative sandwich enzyme immunoassay technique. The microtiter plate has been pre-coated with a monoclonal antibody specific for FDP. Standards or samples are then added to the microtiter plate wells and FDP if present, will bind to the antibody pre-coated wells. In order to quantitatively determine the amount of FDP present in the sample, a standardized preparation of horseradish peroxidase (HRP) -conjugated polyclonal antibody, specific for FDP are added to each well to "sandwich" the FDP immobilized on the plate. The microtiter plate undergoes incubation, and then the wells are thoroughly washed to remove all unbound components. Next, substrate solutions are added to each well. The enzyme (HRP) and substrate are allowed to react over a short incubation period. Only those wells that contain FDP and enzyme-conjugated antibody will exhibit a change in color. The enzyme-substrate reaction is terminated by addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The FDP concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://www.aaabiotech.com/fdp-rabbit-elisa-kits/fibrinogen-degradation-productb74/83950>