

Human Prostaglandin E1 ELISA Kit | PGE1 elisa kit

Human Prostaglandin E1 ELISA Kit

CATALOG #
AAA84129

REACTIVITY
Human


PRODUCT SYNONYMS

Prostaglandin E1; Human Prostaglandin E1 ELISA Kit; PGE1 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Human
Specificity	This assay has high sensitivity and excellent specificity for detection of PGE1. No significant cross-reactivity or interference between PGE1 and analogues was observed. NOTE: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between PGE1 and all the analogues, therefore, cross reaction may still exist in some cases.
Samples	Serum, plasma, cell culture supernatants, body fluid and tissue homogenate
Assay Type	Quantitative Competitive or Sandwich
Sensitivity	1.0pg/mL
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for PGE1 elisa kit	Intended Uses: This PGE1 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Human PGE1. This ELISA kit for research use only, not for therapeutic or test applications! Principle of the Assay: PGE1 ELISA kit applies the quantitative sandwich enzyme immunoassay technique. The microtiter plate has been pre-coated with a monoclonal antibody specific for PGE1. Standards or samples are then added to the microtiter plate wells and PGE1 if present, will bind to the antibody pre-coated wells. In order to quantitatively determine the amount of PGE1 present in the sample, a standardized preparation of horseradish peroxidase (HRP)-conjugated polyclonal antibody, specific for PGE1 are added to each well to "sandwich" the PGE1 immobilized on the plate. The microtiter plate undergoes incubation, and then the wells are thoroughly washed to remove all unbound components. Next, substrate solutions are added to each well. The enzyme (HRP) and substrate are allowed to react over a short incubation period. Only those wells that contain PGE1 and enzyme-conjugated antibody will exhibit a change in color. The enzyme-substrate reaction is terminated by addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The PGE1 concentration in each sample is interpolated from this standard curve.
---	--

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com
Online datasheet: <https://www.aaabiotech.com/pge1-human-elisa-kits/prostaglandin-e1/84129>