

Canine sCD23 ELISA Kit | Scd23 elisa kit

Canine sCD23 ELISA Kit

CATALOG #
AAA91684

REACTIVITY
Canine


PRODUCT SYNONYMS

sCD23; Canine sCD23 ELISA Kit; Scd23 elisa kit

PRODUCT SPECIFICATIONS

Reactivity	Canine
Samples	Serum, plasma, Cell Culture Supernatants, body fluid and tissue homogenate
Assay Type	Sandwich
Detection Range	25-500 U/mL
Sensitivity	1.0ug/mL
Intended Uses	This Scd23 ELISA kit is a 1.5 hour solid-phase ELISA designed for the quantitative determination of Canine Scd23. This ELISA kit for research use only, not for therapeutic applications!
Preparation and Storage	Store all reagents at 2-8 degree C.

IMAGES



Standard Curve (Sample)

ADDITIONAL INFORMATION

Related Product Information for Scd23 elisa kit

Principle of the Assay: Scd23 ELISA kit applies the competitive enzyme immunoassay technique utilizing a monoclonal anti-Scd23 antibody and an Scd23-HRP conjugate. The assay sample and buffer are incubated together with Scd23-HRP conjugate in pre-coated plate for one hour. After the incubation period, the wells are decanted and washed five times. The wells are then incubated with a substrate for HRP enzyme. The product of the enzyme-substrate reaction forms a blue colored complex. Finally, a stop solution is added to stop the reaction, which will then turn the solution yellow. The intensity of color is measured spectrophotometrically at 450nm in a microplate reader. The intensity of the color is inversely proportional to the Scd23 concentration since Scd23 from samples and Scd23-HRP conjugate compete for the anti-Scd23 antibody binding site. Since the number of sites is limited, as more sites are occupied by Scd23 from the sample, fewer sites are left to bind Scd23-HRP conjugate. A standard curve is plotted relating the intensity of the color (O.D.) to the concentration of standards. The Scd23 concentration in each sample is interpolated from this standard curve.

For Research Use Only. Not for use in diagnostic or therapeutic procedures. Small volumes of product may become entrapped in the seal of the vial during shipment and storage; if necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the cap.

AAA Biotech | Tel: +1 (619) 604-9440 | Fax: +1 (619) 604-9441 | contact@aaabiotech.com

Online datasheet: <https://www.aaabiotech.com/canine-elisa-kits/scd23/91684>